



## **Gold Abutment Cylinder Instructions for Use**

### **Indications**

Intra-Lock® UCLA Gold Abutment Cylinders are used with Intra-Lock® dental implants to create custom prosthetics in partially or fully edentulous patients and can be used in the mandible or maxilla. The abutments may be used in single and/or multiple tooth application in the mandible or maxilla.

### **Contraindications**

Patients with alcohol addiction or psychiatric disorders, blood dyscrasias, uncontrolled diabetes, hyperthyroidism, oral infections, malignancies or patients who have had myocardial infarction within the last 12 months.

Patients with systemic diseases that compromise the immune system, such as AIDS, patients on medications that would compromise healing of an implant site, patients with a history of poor or noncompliance to oral hygiene procedures, or patients who cannot maintain oral hygiene procedures if implants are placed.

Tobacco usage, bruxism, nail biting, and pencil biting increases the occurrence of complications and failures.

Additionally, do not use with patients who have a hypersensitivity to one or more of the metals contained in the alloy.

### **Material**

Intra-Lock® UCLA Gold Abutment Cylinders are from high gold alloy in the following composition percentages:

58% Au (gold), 22% Pt (platinum), 19% Pd (palladium), 1% Ir (iron)

*Technical data is available for gold alloy material upon request.*

The dental implants and retention screws are manufactured from titanium alloy (Ti 6Al-4V) to ASTM F136.

### **How Supplied**



Intra-Lock® UCLA Gold Abutment Cylinders are non-sterile and intended for single use only. Abutments should be sterilized prior to use by the instructions provided.

Prior to using Intra-Lock® UCLA Gold Abutment Cylinders, inspect packaging and labeling for integrity. If the device is opened, damaged or contaminated in any way, DO NOT USE.

### **Sterilization**



Intra-Lock® UCLA Gold Abutment Cylinders may be sterilized using a full cycle pre-vacuum steam sterilization at a temperature of 132°C for an exposure time of 3 minutes with 5 minutes of drying time.

### **Processing Information**

Laboratory protocol: Fabricate custom abutments by creating a wax-up over the Intra-Lock® UCLA Gold Abutment Cylinders in preparation for a standard investing technique. Maintain the cylinder at the following stabilization temperature ranges.

Conventional technique alloys: between 650 & 700°C. (1200 & 1292°F)

Ceramic alloys: between 800 & 850°C (1472 & 1562°F) for high gold alloys, between 850 & 900°C (1562 & 1652° F) for Palladium based Precious alloys.

Proceed with casting, bringing the alloy to a temperature 150°C over the liquids.

## Warnings

- Dental implant surgery is a complex dental procedure. Appropriate and adequate training in all phases of implant procedures and proper technique is strongly recommended prior to implant use.
- Improper patient selection, diagnosis, treatment planning or technique can result in implant failure and/or loss of supportive bone.
- Care must be taken if performing electrosurgery around a dental implant. Electrosurgery generates heat, which can be conducted through a metallic implant and cause damage to surrounding tissue or bone.
- The external surface of Intra-Lock® Dental Implants should only come in contact with titanium surfaced instruments.

## Warning

***Caution: Federal law restricts this device to sale by or on the order of a licensed dentist or physician.***



Intra-Lock® International Inc.  
6560 West Rogers Circle, Bldg. 24  
Boca Raton FL 33487 – USA

[www.intra-lock.com](http://www.intra-lock.com)



Intra-Lock System Europa, S.p.A.  
I-84100 Salerno  
+ 39 089 233 045

 0086

IFU-ABGC (EN) (07/2018)